



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L3431-240911)

This is to certify that

Alltestek Co.,LTD
Calibration Laboratory

No.145, Sec.1, Sanmin Rd., Zhongli Dist., Taoyuan City, Taiwan

is accredited in respect of laboratory

- Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018
Accreditation Number : 3431
Originally Accredited : December 01, 2017
Effective Period : December 01, 2023 to November 30, 2026
Accredited Scope : Calibration Field, see described in the Appendix

Yi-Ling Chen



Scan to verify

Yi-Ling Chen
President, Taiwan Accreditation Foundation
September 11, 2024

Accreditation Number : 3431

Laboratory Head : LIAO, Chun-Pin

Electricity

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units	explanation	value	units
KF1001 DC Voltage Source DC Voltage Meter	Agilent/3458A FLUKE/5700A	In-house method: DC Voltage Calibration Procedure (Report No.: ECL-SCP-01)	100	mV	100	mV	DC Voltage Source	16	$\mu\text{V}/\text{V}$
			1	V	1	V	DC Voltage Source	10	$\mu\text{V}/\text{V}$
			10	V	10	V	DC Voltage Source	10	$\mu\text{V}/\text{V}$
			100	V	100	V	DC Voltage Source	12	$\mu\text{V}/\text{V}$
			1000	V	1000	V	DC Voltage Source	14	$\mu\text{V}/\text{V}$
			100	mV	100	mV	DC Voltage Meter	18	$\mu\text{V}/\text{V}$
			1	V	1	V	DC Voltage Meter	10	$\mu\text{V}/\text{V}$
			10	V	10	V	DC Voltage Meter	9	$\mu\text{V}/\text{V}$
			100	V	100	V	DC Voltage Meter	11	$\mu\text{V}/\text{V}$
			1000	V	1000	V	DC Voltage Meter	13	$\mu\text{V}/\text{V}$
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									
KF1002 DC Current Source DC Current Meter	Agilent/3458A FLUKE/5700A	In-house method: DC Current Calibration Procedure (Report No.: ECL-SCP-02)	100	μA	100	μA	DC Current Source	46	$\mu\text{A}/\text{A}$
			1	mA	1	mA	DC Current Source	43	$\mu\text{A}/\text{A}$
			10	mA	10	mA	DC Current Source	43	$\mu\text{A}/\text{A}$
			100	mA	100	mA	DC Current Source	58	$\mu\text{A}/\text{A}$
			1	A	1	A	DC Current Source	0.15	mA/A
			100	μA	100	μA	DC Current Meter	0.16	mA/A
			1	mA	1	mA	DC Current Meter	72	$\mu\text{A}/\text{A}$
			10	mA	10	mA	DC Current Meter	72	$\mu\text{A}/\text{A}$
			100	mA	100	mA	DC Current Meter	83	$\mu\text{A}/\text{A}$
			1	A	1	A	DC Current Meter	0.13	mA/A
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KF1011 AC Voltage Source AC Voltage Meter	Agilent/3458A FLUKE/5700A	In-house method: AC Voltage Calibration Procedure (Report No.: ECL-SCP-03)	100	mV	100	mV	AC Voltage Source @freq.: 60 Hz	0.18	mV/V
			1	V	1	V	AC Voltage Source @freq.: 60 Hz	0.17	mV/V
			10	V	10	V	AC Voltage Source @freq.: 60 Hz	0.17	mV/V
			100	V	100	V	AC Voltage Source @freq.: 60 Hz	0.29	mV/V
			100	mV	100	mV	AC Voltage Source @freq.: 1 kHz	0.23	mV/V
			1	V	1	V	AC Voltage Source @freq.: 1 kHz	0.23	mV/V
			10	V	10	V	AC Voltage Source @freq.: 1 kHz	0.23	mV/V
			100	V	100	V	AC Voltage Source @freq.: 1 kHz	0.29	mV/V
			100	mV	100	mV	AC Voltage Meter @freq.: 60 Hz	0.22	mV/V
			1	V	1	V	AC Voltage Meter @freq.: 60 Hz	97	μV/V
			10	V	10	V	AC Voltage Meter @freq.: 60 Hz	98	μV/V
			100	V	100	V	AC Voltage Meter @freq.: 60 Hz	0.12	mV/V
			100	mV	100	mV	AC Voltage Meter @freq.: 1 kHz	0.22	mV/V
			1	V	1	V	AC Voltage Meter @freq.: 1 kHz	97	μV/V
10	V	10	V	AC Voltage Meter @freq.: 1 kHz	98	μV/V			
100	V	100	V	AC Voltage Meter @freq.: 1 kHz	0.12	mV/V			
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									
KF1012 AC Current Source AC Current Meter	Agilent/3458A FLUKE/5700A	In-house method: AC Current Calibration Procedure (Report No.: ECL-SCP-05)	100	μA	100	μA	AC Current Source @freq.: 60 Hz	1.1	mA/A
			1	mA	1	mA	AC Current Source @freq.: 60 Hz	0.94	mA/A
			10	mA	10	mA	AC Current Source @freq.: 60 Hz	0.94	mA/A
			100	mA	100	mA	AC Current Source @freq.: 60 Hz	0.94	mA/A
			1	A	1	A	AC Current Source @freq.: 60 Hz	1.2	mA/A
			100	μA	100	μA	AC Current Source @freq.: 1 kHz	1.1	mA/A
			1	mA	1	mA	AC Current Source @freq.: 1 kHz	0.60	mA/A
			10	mA	10	mA	AC Current Source @freq.: 1 kHz	0.60	mA/A
			100	mA	100	mA	AC Current Source @freq.: 1 kHz	0.60	mA/A
			1	A	1	A	AC Current Source @freq.: 1 kHz	1.4	mA/A
			100	μA	100	μA	AC Current Meter @freq.: 60 Hz	0.37	mA/A
			1	mA	1	mA	AC Current Meter @freq.: 60 Hz	0.22	mA/A



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KF1012 AC Current Source AC Current Meter	Agilent/3458A FLUKE/5700A	In-house method: AC Current Calibration Procedure (Report No.: ECL-SCP-05)	10	mA	10	mA	AC Current Meter @freq.: 60 Hz	0.22	mA/A
			100	mA	100	mA	AC Current Meter @freq.: 60 Hz	0.22	mA/A
			1	A	1	A	AC Current Meter @freq.: 60 Hz	0.80	mA/A
			100	μA	100	μA	AC Current Meter @freq.: 1 kHz	0.37	mA/A
			1	mA	1	mA	AC Current Meter @freq.: 1 kHz	0.22	mA/A
			10	mA	10	mA	AC Current Meter @freq.: 1 kHz	0.22	mA/A
			100	mA	100	mA	AC Current Meter @freq.: 1 kHz	0.22	mA/A
			1	A	1	A	AC Current Meter @freq.: 1 kHz	0.80	mA/A
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									
KF3001 DC Resistance Source DC Resistance Meter	Agilent/3458A FLUKE/5700A	In-house method: DC Resistance Calibration Procedure (Report No.: ECL-SCP-04)	1	Ω	1	Ω	DC Resistance Source	80	μΩ/Ω
			10	Ω	10	Ω	DC Resistance Source	28	μΩ/Ω
			100	Ω	100	Ω	DC Resistance Source	24	μΩ/Ω
			1	kΩ	1	kΩ	DC Resistance Source	17	μΩ/Ω
			10	kΩ	10	kΩ	DC Resistance Source	16	μΩ/Ω
			100	kΩ	100	kΩ	DC Resistance Source	17	μΩ/Ω
			1	MΩ	1	MΩ	DC Resistance Source	25	μΩ/Ω
			10	MΩ	10	MΩ	DC Resistance Source	75	μΩ/Ω
			1	Ω	1	Ω	DC Resistance Meter	0.12	mΩ/Ω
			10	Ω	10	Ω	DC Resistance Meter	33	μΩ/Ω
			100	Ω	100	Ω	DC Resistance Meter	20	μΩ/Ω
			1	kΩ	1	kΩ	DC Resistance Meter	16	μΩ/Ω
			10	kΩ	10	kΩ	DC Resistance Meter	14	μΩ/Ω
			100	kΩ	100	kΩ	DC Resistance Meter	17	μΩ/Ω
1	MΩ	1	MΩ	DC Resistance Meter	25	μΩ/Ω			
10	MΩ	10	MΩ	DC Resistance Meter	49	μΩ/Ω			
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KF4008 Spectrum Analyzer (On-site calibration included)	rubidium Freq. standard /Fluke /910R Counter /Agilent/53132A signal Generator /Keysight/E8247C power meter/sensor /Agilent/E4418B /Keysight /N8487A/ E9304A /H18 spectrum analyzer /Agilent E4446A	In-house method: spectrum analyzer calibration procedure (Document No.: SPA-SCP-01)	10	MHz	10	MHz	spectrum base freq. accuracy	1.2×10^{-8}	
			250	kHz	500	kHz	central freq. reading value (bandwidth 1 MHz)	9.4×10^{-9}	
			500	kHz	1	MHz	central freq. reading value (bandwidth 1 MHz)	7.9×10^{-9}	
			1	MHz	10	MHz	central freq. reading value (bandwidth 1 MHz)	6.1×10^{-9}	
			10	MHz	40	GHz	central freq. reading value (bandwidth 1 MHz)	5.3×10^{-9}	
			-30	dBm	10	dBm	voltage level (freq. response @0.25~500 MHz)	0.15	dB
			-30	dBm	10	dBm	voltage level (freq. response @0.5~40 GHz)	0.20	dB
Approval Signatory: HSU, Ming-Lu									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		explanation	value
KF4012 Arbitrary function generator (on-site calibration included)	Digital multimeter /HP/Agilent 3458A Counter/HP /Agilent 53132A Power meter /sensor/Agilent E4418B/E9304A Digital oscilloscope /Tektronix DPO7254C Spectrum analyzer /Agilent E4446A	In-house method: arbitrary function generator calibration procedure (document no.: AFG-SCP-01)	0.02	V	0.2	V	Voltage (amplitude)/1 kHz	0.09	%
			0.2	V	2.5	V	Voltage (amplitude)/1 kHz	0.07	%
			2.5	V	10	V	Voltage (amplitude)/1 kHz	0.08	%
			1	mV	100	mV	Voltage (DC Bias)	0.12	%
			0.1	V	1	V	Voltage (DC Bias)	0.08	%
			1	V	10	V	Voltage (DC Bias)	0.08	%
			1	kHz	240	MHz	Frequency	2.5x10 ⁻⁸	
			-20	dBm	10	dBm	Sine wave flatness/(0.1~240) MHz, Reference 100 kHz	0.19	dB
			10	dBm	18	dBm	Sine wave flatness/(0.1~240) MHz, Reference 100 kHz	0.19	dB
			10	dBm	18	dBm	Sine wave flatness/(100~240) MHz, Reference 100 kHz	0.38	dB
			1	MHz	240	MHz	Spurious signal level	0.90	dB
			1	kHz	20	kHz	Total Harmonic Distortion	0.07	%
			2.5	ns	2.5	ns	Rise Time (SQ10MHz)	0.11	ns
			5	ns	5	ns	Rise Time (SQ10MHz)	0.28	ns
13	ns	13	ns	Rise Time (SQ10MHz)	0.53	ns			
Approval Signatory: HSU, Ming-Lu									



Electromagnetics

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1001 Microwave power meter	HP 8478B Keysight 11683	In-house method: PMC-SAP-01	1	mW	1	mW	Power Reference: Frequency 50 MHz	1.2	%
			-25 (3.16)	dBm (μ W)	20 (100)	dBm (mW)	Power Range	0.32	%
Approval Signatory: LIANG, Sung-Chun; HSU, Ming-Lu									
KG1001 Microwave power sensor	HP 8478B Keysight N432A Agilent 11667A	In-house method: PSC-SCP-01	0.8		1		Power 1 mW Frequency 10 MHz to 50 MHz	3.4	%
			0.8		1		Power 1 mW Frequency 50 MHz to 2 GHz	3.4	%
			0.8		1		Power 1 mW Frequency > 2 GHz to 12 GHz	3.6	%
			0.8		1		Power 1 mW Frequency > 12 GHz to 13 GHz	3.8	%
			0.8		1		Power 1 mW Frequency > 13 GHz to 14 GHz	4.0	%
			0.8		1		Power 1 mW Frequency > 14 GHz to 16 GHz	3.8	%
			0.8		1		Power 1 mW Frequency > 16 GHz to 17 GHz	3.6	%
			0.8		1		Power 1 mW Frequency > 17 GHz to 18 GHz	4.6	%
Approval Signatory: LIANG, Sung-Chun; HSU, Ming-Lu									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1002 Short circuit, Opener, Terminator, Mismatch, Attenuator, microwave components	Keysight /85055-60003 Keysight /85055-60004	In-house method: SPC-SCP-01	0.00		<0.10		Reflection coefficient Type N @10 MHz to 500 MHz	0.0064	
			0.10		<0.20		Reflection coefficient Type N @10 MHz to 500 MHz	0.0074	
			0.20		<0.30		Reflection coefficient Type N @10 MHz to 500 MHz	0.0083	
			0.30		<0.40		Reflection coefficient Type N @10 MHz to 500 MHz	0.0093	
			0.40		<0.50		Reflection coefficient Type N @10 MHz to 500 MHz	0.011	
			0.50		<0.60		Reflection coefficient Type N @10 MHz to 500 MHz	0.012	
			0.60		<0.70		Reflection coefficient Type N @10 MHz to 500 MHz	0.014	
			0.70		<0.80		Reflection coefficient Type N @10 MHz to 500 MHz	0.015	
			0.80		<0.90		Reflection coefficient Type N @10 MHz to 500 MHz	0.017	
			0.90		<1.00		Reflection coefficient Type N @10 MHz to 500 MHz	0.019	
			1.00		1.00		Reflection coefficient Type N @10 MHz to 500 MHz	0.020	
			0.00		<0.10		Reflection coefficient Type N @>500 MHz to 2 GHz	0.0064	
			0.10		<0.20		Reflection coefficient Type N @>500 MHz to 2 GHz	0.0074	
			0.20		<0.30		Reflection coefficient Type N @>500 MHz to 2 GHz	0.0083	
			0.30		<0.40		Reflection coefficient Type N @500 MHz to 2 GHz	0.0093	



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1002 Short circuit, Opener, Terminator, Mismatch, Attenuator, microwave components	Keysight /85055-60003 Keysight /85055-60004	In-house method: SPC-SCP-01	0.40		<0.50		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.011	
			0.50		<0.60		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.012	
			0.60		<0.70		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.014	
			0.70		<0.80		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.015	
			0.80		<0.90		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.017	
			0.90		<1.00		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.019	
			1.00		1.00		Reflection coefficient Type N @>500 M Hz to 2 GHz	0.020	
			0.00		<0.10		Reflection coefficient Type N @>2 GHz to 8 GHz	0.013	
			0.10		<0.20		Reflection coefficient Type N @>2 GHz to 8 GHz	0.014	
			0.20		<0.30		Reflection coefficient Type N @>2 GHz to 8 GHz	0.015	
			0.30		<0.40		Reflection coefficient Type N @>2 GHz to 8 GHz	0.017	
			0.40		<0.50		Reflection coefficient Type N @>2 GHz to 8 GHz	0.019	
			0.50		<0.60		Reflection coefficient Type N @>2 GHz to 8 GHz	0.021	
			0.60		<0.70		Reflection coefficient Type N @>2 GHz to 8 GHz	0.024	
			0.70		<0.80		Reflection coefficient Type N @>2 GHz to 8 GHz	0.028	



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1002 Short circuit, Opener, Terminator, Mismatchter, Attenuator, microwave components	Keysight /85055-60003 Keysight /85055-60004	In-house method: SPC-SCP-01	0.80		<0.90		Reflection coefficient Type N @>2 GHz to 8 GHz	0.032	
			0.90		<1.00		Reflection coefficient Type N @>2 GHz to 8 GHz	0.036	
			1.00		1.00		Reflection coefficient Type N @>2 GHz to 8 GHz	0.037	
			0.00		<0.10		Reflection coefficient Type N @>8 GHz to 18 GHz	0.013	
			0.10		<0.20		Reflection coefficient Type N @>8 GHz to 18 GHz	0.015	
			0.20		<0.30		Reflection coefficient Type N @>8 GHz to 18 GHz	0.017	
			0.30		<0.40		Reflection coefficient Type N @>8 GHz to 18 GHz	0.019	
			0.40		<0.50		Reflection coefficient Type N @>8 GHz to 18 GHz	0.023	
			0.50		<0.60		Reflection coefficient Type N @>8 GHz to 18 GHz	0.027	
			0.60		<0.70		Reflection coefficient Type N @>8 GHz to 18 GHz	0.031	
			0.70		<0.80		Reflection coefficient Type N @>8 GHz to 18 GHz	0.037	
			0.80		<0.90		Reflection coefficient Type N @>8 GHz to 18 GHz	0.043	
			0.90		<1.00		Reflection coefficient Type N @>8 GHz to 18 GHz	0.050	
			1.00		1.00		Reflection coefficient Type N @>8 GHz to 18 GHz	0.052	
Approval Signatory: LIANG, Sung-Chun; HSU, Ming-Lu									



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1002 Short circuit, Opener, Terminator, Mismatch, Attenuator, microwave components	Keysight /85055-60003 Keysight /85055-60004	In-house method: SPC-SCP-01	>-10	dB	0	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.063	dB
			>-20	dB	-10	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.071	dB
			>-30	dB	-20	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.086	dB
			>-40	dB	-30	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.10	dB
			>-50	dB	-40	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.12	dB
			>-60	dB	-50	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.14	dB
			-60	dB	-60	dB	Transmission coefficient Type N @10 MHz to 500 MHz	0.19	dB
			>-10	dB	0	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.062	dB
			>-20	dB	-10	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.070	dB
			>-30	dB	-20	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.086	dB
			>-40	dB	-30	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.10	dB
			>-50	dB	-40	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.11	dB
			>-60	dB	-50	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.14	dB
			-60	dB	-60	dB	Transmission coefficient Type N @>500 M Hz to 2 GHz	0.19	dB
			>-10	dB	0	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.13	dB



calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KG1002 Short circuit, Opener, Terminator, Mismatchter, Attenuator, microwave components	Keysight /85055-60003 Keysight /85055-60004	In-house method: SPC-SCP-01	>-20	dB	-10	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.13	dB
			>-30	dB	-20	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.14	dB
			>-40	dB	-30	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.15	dB
			>-50	dB	-40	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.16	dB
			>-60	dB	-50	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.19	dB
			-60	dB	-60	dB	Transmission coefficient Type N @>2 GHz to 8 GHz	0.25	dB
			>-10	dB	0	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.23	dB
			>-20	dB	-10	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.23	dB
			>-30	dB	-20	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.24	dB
			>-40	dB	-30	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.26	dB
			>-50	dB	-40	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.27	dB
			>-60	dB	-50	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.29	dB
			-60	dB	-60	dB	Transmission coefficient Type N @>8 GHz to 18 GHz	0.36	dB
Approval Signatory: LIANG, Sung-Chun; HSU, Ming-Lu									



Time And Frequency

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KJ0200 1. Various types of signal generator 2. Various frequency counters	1. Controlled Frequency Standards /FLUKE 910R 2. Universal Counter /Agilent 53132A	In-house method: Time and frequency calibration procedures (Document No.: TFC-SCP-01)	10	MHz	10	MHz		1.5×10^{-8}	
Approval Signatory: LIN, Shun-Hui; LIANG, Sung-Chun									

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

